

## Clonazepam Level

Order Name: **CLONAZEPA**  
Test Number: 4001540  
Revision Date: 12/12/2022

TEST NAME	METHODOLOGY	LOINC CODE
Clonazepam Level	High Performance Liquid Chromatography	3494-2

### SPECIMEN REQUIREMENTS

Specimen	Specimen Volume (min)	Specimen Type	Specimen Container	Transport Environment
Preferred	4 mL (1.1)	Plasma	EDTA (Lavender Top)	Refrigerated
Alternate 1	4 mL (1.1)	Serum	Clot Activator (Red Top, No-Gel)	Refrigerated
Alternate 2	1 mL (0.1)	Plasma	Sodium Heparin (Green Top / No-Gel)	Refrigerated

**Instructions**

**Specimen Type:** Red-top tube, lavender-top (EDTA) tube OR green-top (heparin) tube. **DO NOT USE A GEL-BARRIER TUBE.** The use of gel-barrier tubes is not recommended due to slow absorption of the drug by the gel. Depending on the specimen volume and storage time, the decrease in drug level due to absorption may be clinically significant.

**Specimen Storage:** Refrigerate

**Specimen Collection:** Transfer separated serum or plasma to a plastic transport tube. Collect specimen immediately prior to next dose unless specified otherwise.

**Specimen Stability:** Ambient: 14 days, Refrigerated : 14 days, Frozen: 14 days

### GENERAL INFORMATION

Expected TAT	2 - 4 days working day.
Performing Labcorp Test Code	071712
Notes	Labcorp Test Code: 071712
CPT Code(s)	80346 (Alt code: G0480)
Lab Section	Reference Lab