

Thyroglobulin Panel

Order Name: **Thyroglob Pnl**

Test Number: 3612483

Revision Date: 12/22/2023

TEST NAME	METHODOLOGY	LOINC CODE
Thyroglobulin Autoantibody (TG Ab)	Enzyme Immunoassay	8098-6
Thyroglobulin Level	Immunoassay (IA)	3013-0

SPECIMEN REQUIREMENTS

Specimen	Specimen Volume (mL)	Specimen Type	Specimen Container	Transport Environment
Preferred	3 mL (1.8 mL)	Serum	Clot Activator SST	Frozen

Instructions

Preferred is Two Aliquots of Serum:
(1) Allow specimen to clot fully, centrifuge, remove 2mL of Serum and Freeze.
(2) 1 mL Serum kept Refrigerated.
Serum stability: Ambient 8 hours, Refrigerated 3 days, Frozen 2 months

Thyroglobulin by immunoassay is highly susceptible to biotin interference. Patients taking biotin supplements – particularly at high doses – may show falsely low or undetectable thyroglobulin levels. Please consult with lab if biotin interference is suspected.

GENERAL INFORMATION

Testing Schedule	Mon-Fri
Expected TAT	1-4 days from set up
CPT Code(s)	86800, 84432
Lab Section	Chemistry